

***Supply of items (pH meter with micro probe) Tender Enquiry No.6-15/ST/DBT/AB/2015-16 dated 14.09.2015 Last Date-08.10.2015 (under this category Tender Form may be downloaded from the top of this page)***

**Annexure – I**

### **LIST OF REQUIREMENTS**

#### **pH meter with micro probe-One Unit**

- Refillable double junction glass electrode
- pH range-2.00 to 20.00
- Dual-line LCD display
- ATC probe
- Electrode stand & Electrode fill solution
- Current 100/240 Volt
- Automatic temperature compensation (0-100 °C)
- High pH resolution (up to 0.001)
- pH accuracy ( $\pm 0.002$ ); Temperature accuracy 0.5 °F
- Data logger up to 500
- Three years Warranty from date of installation
- F.O.R. CARI, Izatnagar, Bareilly

**Note:** - The envelope shall be addressed to the Senior Administrative Officer (Stores), ICAR-Central Avian Research Institute, Izatnagar-243122, Bareilly, UP and it should also bear the “Tender Enquiry no. and the words “do not open before 15 hrs on 08.10.2015 (The time and date of opening of the tenders)” . The deadline of the submission of bids is 08.10.2015 upto 02:00 PM.

Please specify the make and model for above items and also attach relevant document/literature (if available) illustrating detail specifications of the product.

Special instruction (if any)\_\_\_\_\_

**TERMS AND CONDTITIONS**

1. Submit your Tender along with the prescribed tender form (LVP) available on the Institute website. The sealed envelope shall be addressed to the Senior Administrative Officer (Stores), ICAR-Central Avian Research Institute, Izatnagar-243122, Bareilly and super scribed “ **Tender Enquiry Number and Name of the Equipment/goods etc. at the top of the envelope and the word “DO NOT OPEN BEFORE 15 hours on 08.10.2015** (the time and date of opening of the tenders).” The deadline of the submission of bids is 08.10.2015 upto 02:00 PM.
2. The tender forms without the required amount of Bid security and late tenders will not accepted in any circumstance and liable to be rejected.
3. Conditional and incomplete tenders shall not be entertained in any circumstances and liable to be rejected. The following are the important points, for which a tender may be declared as unresponsive and to be ignored, during the initial scrutiny.
  - (i) The tender is unsigned.
  - (ii) The tenderer is not eligible (Example: The tender enquiry condition says that the bidder has to be a registered SSI unit; but the tenderer is a say, Large Scale Unit)
  - (iii) The tender validity is shorter than the reputed period.
  - (iv) The tenderer has quoted for goods manufactured by a different firm without the required authority letter from the proposed manufacturer.
  - (v) Tenderer has not agreed to give the required performance security.
  - (vi) The goods quoted are sun-standard, not meeting the required specifications etc.
  - (vii) Against a schedule in the List of Requirement (incorporated in the tender enquiry, the tenderer has not quoted for the entire requirement as specified in that schedule. (Example: In a schedule, it has been stipulated that the tenderer will supply the equipments, install and commission it and also train the purchaser’s operators for operating the equipment. The tenderer has however, quoted only for supply of the equipment).
  - (viii) The tenderer has not agreed to some essential condition(s) incorporated in the tender enquiry. (Example: Some such important essential conditions are-terms of payment, liquidated damages clause, warranty clause, dispute resolution mechanism, applicable law and any other important condition having significant bearing on the cost/utility/performance of the required goods, etc.
4. Only original and standard product(s) of reputed firm(s)/make with company sealed packing will be accepted only.
5. The Indian Agency commission (if any) should be shown separately in the schedule of rates.
6. Installation to be done within 10 days from the date of Issue of our letter for this purpose, failing which liquidation damage clause shall apply.
7. Guarantee/Warranty of the equipment should be 36 month from the date of installation.

8. Charges of the Annual maintenance for next 05 years after expiring of guarantee/warranty period must be mentioned.
9. The successful tenderers will have to deposit the Performance security @05% of the quoted price for the guarantee/warranty period of the equipment/item within 21 days from the date of issue of purchase order, failing which the order will be treated as cancelled and the bid security deposited for the purpose will be forfeited.
10. The validity of tender offer should be at the minimum for 90 days from the date of opening of tender. The quantity/number of equipment/item may be increased or decreased as the case may be.
11. Payment may be made only after successful supply/installation of the equipment/item in the shape of demand draft/sight draft/RTGS only. Complete banking must be enclosed with the tender documents.
12. Director, CARI reserves the right to accept or reject any one or all tender(s) without assigning any reason thereof and his decision will be final in this regard.
13. Performa Invoice of Principal manufacturer, if any, must be attached alongwith tender.
14. Detailed specifications of the equipments/goods to be supplied, only one make/model be quoted and no alternative make/model will be considered.
15. Product Literature/Leaflets should be essentially attached with the tender document.
16. Certified copy of PAN/ITCC issued by Dept. Of Income Tax.
17. Certified copy of Tin Number/CST registration/clearance certificate.
18. Illustrated Catalogues of the equipment to be submitted with the tender.